

# Sienco, Inc.

5721 Arapahoe Ave, Unit A1-A

Boulder, CO 80303 USA

800/432-1624 303/420-1148 303/379-4403 (Fax)

http://www.sienco.com email: sienco@sienco.com

## SonACT™ Kit

#800-0432 (100s), #800-0431 (24s)

For In Vitro Diagnostic Use

### INTENDED USE

The SonACT Kit is an in vitro diagnostic reagent for use with the Sonoclot® Coagulation & Platelet Function Analyzer System. The SonACT test is a celite activated whole blood clotting time test. It may also be used with citrated whole blood and plasma.

The SonACT Kit is intended for general purpose global hemostasis monitoring including: clot detection, fibrin formation, platelet function, and hyperfibrinolysis. The monitoring information is typically used for anticoagulant management up to high heparin levels (0 to approximately 6 units per ml), hypercoagulable and/or hypocoagulable screening, platelet function assessment, and hyperfibrinolysis screening.

When used with the Sonoclot Analyzer System, the SonACT test provides quantitative Onset/ACT and Clot Rate results as well as qualitative and quantitative information on the Sonoclot Signature including: fibrin formation, platelet function, and hyperfibrinolysis.

### SUMMARY AND PRINCIPLES

The activated clotting time is the amount of time it takes to form a clot by contact activation of the coagulation cascade. A variety of materials are commonly used in ACT tests for contact activation, including glass beads (silica), celite (diatomaceous earth) and clay (kaolin). Each of these activators has unique activation characteristics and therefore different activators are more desirable than others in different applications. Celite activation is highly effective for high dose heparin management. A celite test, however, can be significantly prolonged by the protease inhibitor aprotinin. This may lead to inadequate anticoagulation during bypass. For many people, a kaolin test shows less aprotinin effect than the celite test, but these results are highly variable from patient to patient. Sienco's aprotinin insensitive aiACT test is less affected by aprotinin than any celite or kaolin activated test examined by the manufacturer. If the SonACT test is used for heparin management in the presence of aprotinin, the minimum ACT threshold should be increased based on the amount of aprotinin used. For aprotinin levels less than 300 KIU/ml, the ACT target should be extended to at least 700 seconds.

### REAGENTS

Each SonACT Kit contains lidded colorless plastic activation cuvettes, probes and these instructions for use. The activation cuvettes contain a controlled amount of celite and a magnetic stir bar. Packaging sizes of 100 (800-0432) and 24 (800-0431) are available.

### STORAGE

Store at room temperature. Protect from heat.

### PROCEDURE

#### A. Equipment Required:

1. Sonoclot Analyzer System

#### B. Preparation

1. Make sure that the Sonoclot Analyzer is turned on and warmed up with the head assembly in the down position. Check that the Printer is ON and ON-LINE.
2. Sharply tap the cuvette against a hard surface, cap side up, to dislodge any activation powder from the sides and lid of the cuvette. It is normal for some of the powder to remain on the sides of the cuvette.
3. Place the cuvette into the instrument warming well so it will be warm and ready to go when the sample is ready for analysis. Allow the cuvette to warm for at least 5 minutes prior to testing. Probes fit into the cuvette lids for convenient storage. Several cuvettes may be prepared and placed in the warming wells prior to initial testing.

#### C. Cuvette and Probe Set-up

1. Proper mounting of the probe on to the probe mount hub is the most important step in ensuring the reliable performance of the Sonoclot Analyzer.
2. Open the head assembly by tilting it backwards on its hinge.
3. With a slight twisting motion, push a clean disposable tubular probe over the probe mount hub inside the head assembly. This motion should result in the probe sliding straight over the probe mount hub. The hub should not move sideways. The probe must be fully seated on the probe mount hub for proper operation.
4. WHILE THE CUVETTE IS STILL IN THE WARMING WELL, remove the cuvette lid by popping it off with your thumb and forefinger. Do not remove the cuvette lid while the cuvette is in the cuvette holder; the cuvette holder may break. Insert the opened cuvette into the cuvette holder with a slight twisting motion. Ensure that the cuvette is fully seated in the cuvette holder. Check that the cuvette contains a stir bar.

#### D. Obtaining the Blood Sample


Native whole blood must be analyzed within 2 minutes or less from collection. Please observe the following precautions when drawing the blood sample:

- Sample withdrawal must be smooth, slow and atraumatic. Under no conditions should a sample be drawn with force. The sensitivity of platelets to disturbance makes good sampling technique especially important.
- Care should be exercised in deciding where the sample will be drawn to prevent possible heparin contamination.
- A two syringe technique is recommended for drawing the blood sample. The first syringe of 2 to 3 ml is discarded and the second syringe is used for the sample. Plastic syringes are mandatory to avoid uncontrolled glass activation.

Citrated whole blood or plasma samples can also be used with the SonACT test. Please refer to the Operator's Manual for specific sample handling and testing instructions.

#### E. Running the Sonoclot Analyzer

1. For native whole blood testing, fill the warmed cuvette with the blood sample from the syringe so that the fluid level is slightly below the inner rim of the cuvette as shown below (≈360 µl). Transfer the sample from the syringe into the cuvette either with or without a blunt needle or cannula. Immediately press the START/STOP switch to the START position. The magnetic stirrer will automatically rotate and the Printer will begin to print.
2. After 10 seconds, the Sonoclot Analyzer will beep and the panel display will prompt you to CLOSE HEAD. Close the head assembly. At this time, if you wish to run an analysis for more than twenty to thirty minutes, carefully place a drop of SonOil™ on the top of the sample. The SonOil will prevent the clot from drying out and forming a crust on the top of the sample.
3. The sample is initially a liquid. After several minutes, the sample begins to evolve into a clot. The instrument detects this initial clot formation, beeps and displays the time that the sample remained a liquid above the ACT/ONSET legend.



- During the next several minutes of analysis, the fibrinogen converts into a fibrin gel. The Sonoclot Analyzer determines this rate of conversion by calculating the rate of change in the Clot Signal Value and reports this rate above the Clot RATE legend.
4. Continue to allow the instrument to run in order to obtain information on platelet function and fibrinolysis. If you are interested in monitoring clot retraction (platelet function), you should allow the analysis to continue for 20 to 30 minutes or until clot retraction completes.
  5. When your analysis is complete, momentarily press the START/STOP switch to the STOP position to stop the Printer. If you forget press the STOP switch, the Printer will automatically stop after 60 minutes (default value). The automatic shut-off feature can be customized to your specific requirements. Please refer to the Operator's Manual for complete instructions.
  6. Open the head assembly. Remove the tubular probe with a probe extractor or other tool. Use a straight motion that pulls the probe straight away from the probe mount hub. Avoid moving the hub sideways. Properly discard the probe and cuvette. Lower the head assembly to maintain temperature control of the instrument.

## F. Quality Control

SonACT test performance should be verified with reference plasma coagulation controls. If plasmas are being run to QC both the activation cuvettes and Sonoclot Analyzer, testing should be performed once each day prior to sample testing. If plasmas are being run to QC activation cuvettes only, testing should be performed prior to the use of a new shipment and monthly throughout use of the stock. In either case, more frequent testing may be required to comply with local, state and federal QC requirements. Sienco offers the Reference Plasma Quality Control Kit, part # 900-1318 to verify the performance of activation cuvettes and the Sonoclot Analyzer. Sienco offers the Reference Viscosity Oil Quality Control Kit, part # 900-1302 to verify performance of the Sonoclot Analyzer.


## EXPECTED VALUES

Normal results for the SonACT test are summarized in the following table.

SonACT Test (#800-0432 800-0431)	
Native Whole Blood- Normal Population, No Heparin	
Result	Normal Range
ACT/Onset	85-145 seconds
Clot RATE	15-45 Clot Signal Units / minute
Time to Peak	< 30 minutes


It is important to know that normal ranges for a healthy population may be different than average results for specific patient populations. Medication and variations in sampling and operator technique can alter average results at each institution. It is also true that the hemostatic system responds to the stress of surgery in ways that affect the Sonoclot Signature - typically in accelerating clot retraction. Even though there may be sources of variability, the following ranges provide a useful base line.

Standard SonACT, Normal A,  
Native Whole Blood



Onset/ACT: 89  
Clot RATE: 22  
Time to Peak: 17.5


Standard SonACT, Normal B,  
Native Whole Blood



Onset/ACT: 89  
Clot RATE: 22  
Time to Peak: 17.5


## PLATELET FUNCTION TESTING

Coagulation Phases of the Sonoclot® Signature




Platelet function information is obtained during the clot retraction phase of the Sonoclot Signature. As platelets cause the clot to retract, the Sonoclot Analyzer observes changes in the viscoelastic measurement of the clot. Rapid or strong changes indicate greater platelet function. The reference Signature below shows good platelet function because the Sonoclot Signature shows significant changes during the clot retraction phase.

Good Platelet Function



Poor Platelet Function



## OPERATIONAL PRECAUTIONS AND LIMITATIONS

Test result quality depends on proper technique. Carefully follow these precautions.

- 1) Only properly trained lab personnel and healthcare professionals should operate the analyzer.
- 2) Diagnosis should not be based solely on Sonoclot test results. The attending physician is responsible for interpreting the analyzer test results in conjunction with the patient's condition, other test results, and clinical observations.
- 3) Handle materials with care:
  - When taking a blood sample, avoid heparin contamination from catheters.
  - Never use the first sample from a new line to avoid sample contamination with tissue thromboplastin.
  - Keep blood, dirt, and other substances away from the probe mount hub to avoid contaminating the electromechanical transducer.
  - Never reuse a probe or cuvette to prevent thrombin contamination.
- 4) If the analyzer is not at the desired temperature (normally 37 °C) then the analyzer will display an error message and not run the test.
- 5) For consistent results the cuvettes should be warmed prior to analysis. Do not store the cuvettes in the warming wells for extended periods of time (i.e. overnight) to avoid sample degradation from prolonged exposure to heat.
- 6) Always remove the cuvette cap before placing it into the cuvette holder. Failure to do so can damage the transducer. When placing the cuvette in the cuvette holder, verify that the cuvette contains a stir bar.
- 7) The cuvette must be fully seated in the cuvette holder to avoid interference between the probe and stir-bar.
- 8) Do not overfill the cuvette. The proper fill level is 330 - 360µl, slightly below the inner rim of the cuvette.
- 9) Always insert and remove the probe by moving it vertically over the probe mount hub. Never move the hub horizontally. Make sure the probe is fully seated to avoid interference with the stir-bar.
- 10) Perform QC testing to verify proper operation of the analyzer and activation cuvettes.
- 11) Use proper biohazard handling techniques to dispose of probes and cuvettes.
- 12) On rare occasions, mechanical disturbances may cause incorrect results. Always inspect results to ensure that they are consistent.
- 13) Extremely high viscosity blood samples, (immersion response on analyzer > 25 clot signal units) can stratify. Use an external device to mix the blood sample before filling the cuvette.

## PERFORMANCE

The Onset time CV should not exceed 10% for a test sample run side by side on multiple Sonoclot Analyzer Systems. The Clot Rate CV may be slightly greater since the Clot Rate parameter is more technique dependent. Due to sample aging effects, it is not practical to run a test sample multiple times on one Sonoclot Analyzer System to determine a CV. Each trained user should establish a mean and standard deviation on a month to month basis as a quality control monitor.

## BIBLIOGRAPHY

Sonoclot Coagulation & Platelet Function Analyzer Operator's Manual